

2047900-1048 SecurityGuide (2).pdf (SECURED) - Adobe Reader

File Edit View Window Help

45 / 90

78.9%

Comment

Sign

Bookmarks

Introduction

Access Control

Security

Network Information

Anti-Virus Software


Product Vulnerability and Security Patches

Appendix A Audit Logging

Appendix B Passwords

Appendix C Validated Configuration Information

Validated Anti-Virus Software



**WARNING: SYSTEM INSTABILITY**

Do not install or use unvalidated anti-virus software (including unvalidated versions). Doing so may result in system instability or failure. Use only validated anti-virus software in the appropriate language version.

**NOTE:** If the language specific anti-virus software is not available, install the English version of anti-virus software.

The Mac-Lab/Cardiolab version 6.9 systems have been validated to run with the software listed in the following table.

Supported Anti-Virus Software	Supported Languages	Supported Anti-Virus Software Version
CA eTrust Anti-Virus	English, French, German, Italian, Spanish, Chinese and Japanese	18.1.655
McAfee VirusScan Enterprise	English, French, German, Italian, Spanish, Swedish, Dutch, Chinese and Japanese	8.7i, Patch 3
McAfee ePolicy Orchestrator (with McAfee VirusScan Enterprise 8.7i, Patch 3)	English, French, German, Spanish, Chinese and Japanese	v4.5, Patch 1
Symantec EndPoint Protection	English, French, German, Italian, Spanish, Chinese and Japanese	11.0.0
Trend Micro OfficeScan Client/Server Edition	English, French, German, Spanish, Chinese and Japanese	10.0 + SP1

Click on Sign to add text and place signature on a PDF File.

Product Security Database - Windows Internet Explorer

File Edit View Favorites Tools Help

http://prodsecdb.gehealthcare.com/prodsecdb/productAndVulnerMapSearch.do?invoke=search&productKeyword

Product Security Database

### Product - Patch Mapping Search Result

Search By : Product Name/Keyword = Centricity Cardiology INW Server, Product Version = 6.9

17 Items found, displaying all items.

Product Name/Version	Patch Identifier	Patch Assessment Status	Patch Assessment Date	Patch Assessment Details
Centricity Cardiology INW Server6.9	MS10-046	Patch Approved	16-August-2010	
Centricity Cardiology INW Server6.9	MS10-054	Patch Approved	16-August-2010	
Centricity Cardiology INW Server6.9	MS10-061	Patch Approved	27-September-2010	
Centricity Cardiology INW Server6.9	MS13-049	Evaluation In Progress	01-July-2013	
Centricity Cardiology INW Server6.9.5	MS13-049	Evaluation In Progress	01-July-2013	
Centricity Cardiology INW Server6.9.5	MS13-001	Evaluation In Progress	01-July-2013	
Centricity Cardiology INW Server6.9	MS13-001	Evaluation In Progress	04-February-2013	
Centricity Cardiology INW Server6.9	MS13-018	Evaluation In Progress	30-April-2013	
Centricity Cardiology INW Server6.9.5	MS13-018	Evaluation In Progress	01-July-2013	
Centricity Cardiology INW Server6.9	MS12-054	Patch Approved	31-August-2012	
Centricity Cardiology INW Server6.9	MS12-036	Patch Approved	28-June-2012	
Centricity Cardiology INW Server6.9	MS12-017	Patch Approved	22-March-2012	
Centricity Cardiology INW Server6.9	MS12-020	Patch Approved	22-March-2012	
Centricity Cardiology INW Server6.9	MS12-004	Patch Approved	03-February-2012	
Centricity Cardiology INW Server6.9	MS11-053	Patch Approved	08-September-2011	Domain Controller Only>>>
Centricity Cardiology INW Server6.9	MS11-020	Patch Approved	22-April-2011	
Centricity Cardiology INW Server6.9	MS11-015	Patch Approved	22-April-2011	

[Contact Us](#) | [Privacy Policy](#) | [Terms & Conditions](#)

© 2007 General Electric Company  
Build 1.4

Internet | Protected Mode: Off 100%